

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska

APR 14 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Datamaster cdm S/N <u>130492</u>	
Name <u>Shayne Carr</u>	ID# <u>4559</u> Date <u>3-12-10</u>
A Agency <u>AST</u>	Phone # <u>733-2256</u>
Instrument Location <u>Willow</u>	
Alco S/N <u>X301581</u> Target Value <u>.080</u> High Pressure <u>950 psi</u>	
B Alco Test Value Average <u>.082</u> <u>.082</u>	
1 st Alco 2 nd Alco	
Signature <u>[Signature]</u>	
(OVER)	

bmb
4/15/2010

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15th day of April, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130492

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

950PSI

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492 ✓

MARCH 12, 2010 ✓

OPERATOR'S NAME:

CALT SHAYNE M

OPERATOR'S NUMBER: 4559

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

V

O.L. #: V

DEPT/AGENCY: TALO

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .002 ✓

ALCO S/N: X301581 ✓

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 29.42 in ✓

ALCO TARGET .000 11:15

BLANK TEST .000 11:16

INTERNAL STANDARD VERIFIED 11:16

ALCO TV 29.42 in .002 ✓ 11:16

BLANK TEST .000 11:17

SUBJECT SAMPLE .000 11:17

BLANK TEST .000 11:18

ALCO TV 29.42 in .002 ✓ 11:18

BLANK TEST .000 11:19

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492 ✓

MARCH 12, 2010 ✓

TIME 11:13

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.42 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~